

WL 19 Con

AA

LOT

Rev.: 003

WL 19 Con AA 3 niveles

PARAMETRO	WIENER LAB			WIENER LAB			WIENER LAB		
	COUNTER 19			COUNTER 19 CP			COUNTER XS 20		
	NIVEL 1	NIVEL 2	NIVEL 3	NIVEL 1	NIVEL 2	NIVEL 3	NIVEL 1	NIVEL 2	NIVEL 3
WBC (x 10 ⁹ /L)	2,1±0,5	8,2±1,0	21,2±2,5	2,1±0,5	8,2±1,0	21,8±2,5	2,2±0,5	7,9±1,0	21,1±2,5
LYMPH # (x 10 ⁹ /L)	1,2±0,3	2,4±0,7	3,1±1,3	1,2±0,3	2,4±0,7	2,9±1,3	1,4±1,3	2,7±0,7	3,2±1,8
MID # (x 10 ⁹ /L)	0,2±0,2	0,7±0,7	1,3±1,1	0,2±0,2	0,7±0,7	1,2±1,1	0,1±0,1	0,4±0,4	0,8±0,6
GRAN # (x 10 ⁹ /L)	0,7±0,3	5,1±0,7	16,9±1,8	0,7±0,3	5,2±0,7	17,7±1,8	0,7±0,3	4,8±0,7	17,1±1,9
LYMPH % (%)	55,0±12,0	29,5±8,0	14,5±6,0	56,0±12,0	29,0±8,0	13,5±6,0	64,0±12,0	34,0±8,0	15,0±8,0
MID % (%)	11,0±9,0	8,0±7,0	6,0±5,0	9,5±8,0	8,0±7,0	5,5±5,0	6,0±5,0	5,0±4,0	4,0±3,0
GRAN % (%)	34,0±10,0	62,5±8,0	79,5±8,0	34,5±10,0	63,0±8,0	81,0±8,0	30,0±10,0	61,0±8,0	81,0±9,0
HGB (g/dL)	6,1±0,6	13,8±0,9	18,68±1,3	6,1±0,6	13,9±0,9	19,0±1,3	6,1±0,6	13,9±0,9	19,0±1,3
RBC (x 10 ¹² /L)	2,45±0,18	4,85±0,24	5,87±0,30	2,43±0,18	4,81±0,24	5,87±0,30	2,35±0,18	4,59±0,24	5,59±0,30
HCT (%)	18,4±1,5	41,5±2,0	55,8±2,4	18,2±1,5	41,1±2,0	56,1±2,4	17,5±1,5	39,2±2,0	52,5±2,4
MCV (fL)	75,0±5,0	85,5±5,0	95,0±5,0	75,0±5,0	85,5±5,0	95,5±5,0	74,5±5,0	85,5±5,0	94,0±5,0
MCH (pg)	24,9±2,7	28,5±2,5	32,0±2,5	25,1±2,7	28,7±2,5	32,4±2,5	26,0±2,7	30,3±2,5	34,0±2,5
MCHC (g/dL)	33,2±3,0	33,3±3,0	33,7±3,0	33,5±3,0	33,6±3,0	33,9±3,0	34,8±3,0	35,4±3,0	36,2±3,0
RDW-CV (%)	16,5±3,0	15,5±3,0	15,5±3,0	14,5±3,0	14,0±3,0	13,5±3,0	16,5±3,0	15,5±3,0	15,0±3,0
RDW-SD (fL)	36,5±6,0	39,0±6,0	41,5±8,0	35,5±6,0	39,0±6,0	41,5±8,0	44,0±6,0	47,0±6,0	49,5±8,0
PLT (x 10 ⁹ /L)	80±20	260±40	515±60	75±20	255±40	527±60	61±20	255±40	545±60
MPV (fL)	8,8±3,0	8,3±3,0	8,4±3,0	8,5±3,0	7,8±3,0	7,9±3,0	9,2±3,0	9,0±3,0	9,5±3,0

WL 19 Con AA nivel 2

PARAMETRO	NIVEL 2		
	WIENER LAB COUNTER 19	WIENER LAB COUNTER 19 CP	WIENER LAB COUNTER XS 20
WBC (x 10 ⁹ /L)	8,2 ± 1,0	8,2± 1,0	7,9± 1,0
LYMPH # (x 10 ⁹ /L)	2,4 ± 0,7	2,4 ± 0,7	2,7 ± 0,7
MID # (x 10 ⁹ /L)	0,7 ± 0,7	0,7 ± 0,7	0,4 ± 0,4
GRAN # (x 10 ⁹ /L)	5,1 ± 0,7	5,2 ± 0,7	4,8 ± 0,7
LYMPH % (%)	29,5 ± 8,0	29,0 ± 8,0	34,0 ± 8,0
MID % (%)	8,0 ± 7,0	8,0 ± 7,0	5,0 ± 4,0
GRAN % (%)	62,5 ± 8,0	63,0 ± 8,0	61,0 ± 8,0
HGB (g/dL)	13,8 ± 0,9	13,9 ± 0,9	13,9 ± 0,9
RBC (x 10 ¹² /L)	4,85 ± 0,24	4,81 ± 0,24	4,59 ± 0,24
HCT (%)	41,5 ± 2,0	41,1 ± 2,0	39,2 ± 2,0
MCV (fL)	85,5 ± 5,0	85,5 ± 5,0	85,5 ± 5,0
MCH (pg)	28,5 ± 2,5	28,7 ± 2,5	30,3 ± 2,5
MCHC (g/dL)	33,3 ± 3,0	33,6 ± 3,0	35,4 ± 3,0
RDW-CV (%)	15,5 ± 3,0	14,0 ± 3,0	15,5 ± 3,0
RDW-SD (fL)	39,0 ± 6,0	39,0 ± 6,0	47,0 ± 6,0
PLT (x 10 ⁹ /L)	260 ± 40	255 ± 40	255 ± 40
MPV (fL)	8,3 ± 3,0	7,8 ± 3,0	9,0 ± 3,0

Glosario/Glossário/Glossary/Glosariusz

Español	Português	English	Polski
fecha	data	date	data
parámetro	parametro	parameter	parametry
nivel	nível	level	poziom



Wiener lab.